

Certificate of Analysis

| | | | |
|-----------------------|--|---|-----------------|
| Product Name | Recombinant Human Tissue Factor | Batch No. | 20250428 |
| Cat. No. | Y04301 | Storage Conditions | ≤-20℃ |
| Manufacturing Date | 2025.04.28 | Expiration Date | 2028.04.27 |
| Testing Basis | Specification for Recombinant Human Tissue Factor | | |
| Testing Parameters | Quality Specification | Testing Result | Result Judgment |
| Appearance | White powder, Clear colorless liquid after resolution | White powder, Clear colorless liquid after resolution | Compliance |
| pH | 7.0 - 8.0 | 7.4 | Compliance |
| Protein Concentration | 9.0 - 11.0 μg | 9.9 μg | Compliance |
| Hydration | ≤2.5% | 1.6% | Compliance |
| Purity (SDS-PAGE) | ≥95.0% | 97.6% | Compliance |
| Relative Bioactivity | The relative biological activity should be 70%-130% comparison to standard reference | 102% | Compliance |



Jiangsu East-Mab Biomedical Technology Co.,Ltd.

| | | | |
|---------------------|---------------------------------|----------------|-----------------|
| Product Name | Recombinant Human Tissue Factor | Batch No. | 20250428 |
| Testing Parameters | Quality Specification | Testing Result | Result Judgment |
| Bacterial Endotoxin | ≤10.0 EU/mg | ≤10.0 EU/mg | Compliance |
| Sterility Test | Compliance | Compliance | Compliance |
| Conclusion | Compliance | | |



Reported by / Date:

Xiao qing Liu 2025.05.20

Checked by / Date:

Jing meng 2025.05.20

Approved by QA / Date:

Fanning zhu 2025.05.20

