



User Manual

Disclaimer: Products are intended for research use only

SHENTEK **Biofargo**

🌐 www.biofargo.com
📞 804-529-2296
✉️ contact@biofargo.com

Vesivirus 2117 Positive Control

Product No.: 1506743-R01

(For Research Use Only)

■ Product description

SHENTEK® Vesivirus 2117 Positive Control is intended for use with SHENTEK® Virus DNA & RNA Extraction and Detection Kits, and serves as a positive control to monitor the entire process of Vesivirus 2117 RNA extraction and RT-PCR detection assay.

■ Kit contents and storage

Reagent	Quantity	Storage
Vesivirus 2117 Positive Control	100 µL × 3 tubes	below -65°C

Vesivirus 2117 Positive Control can be stored at appropriate conditions for up to 12 months. Please check the expiration date on the labels.

■ Attentions

1. Thaw one tube completely at 2-8°C or on ice, then vortex and spin briefly.
2. Follow the positive control sample (PCS) preparation protocol outlined in the user guide of SHENTEK® Vesivirus 2117 Detection Kit.
3. This product is ready for direct use without dilution.
4. Use only after it is equilibrated to room temperature, and avoid repeated freeze-thaw cycles.
5. Do not mix reagents from different batches.
6. Do not use reagents beyond their expiration date.
7. This product is intended for research use only and is not used for clinical diagnosis or treatment. Use with appropriate caution.
8. As an armored RNA product, this material should be handled in a biosafety cabinet. Experimental waste generated during its use must be sterilized via autoclaving and disposed of in compliance with medical waste regulations.

■ Related products

Product Name	Product No.	Quantity
SHENTEK® Virus DNA & RNA Extraction Kit	1506730	50 Extractions
SHENTEK® Vesivirus 2117 Detection Kit	1506743	50 Reactions

Effective date: 01 Dec. 2024

Support & Contact**SHENTEK**

Biofargo, Inc.

www.biofargo.com

Address: 1716 E Parham Rd Richmond, Va, 23228, USA

E-mail: contact@biofargo.com

Phone: 804-529-2296